



Scope Electric

Manufacturer of professional milk analyzers

MilkoScope

Julie C3 Automatic

Julie C3 Automatic is the successor of **Julie C3**.

Case – Aluminum 3 mm

LCD display, 4 rows, 16 symbols

Scope – system for automatic start of the analysis

Built-in device for coping with distortions, caused by electricity malfunctions.

System for control of the sound signal.

PC connection option with a cable and software, included in the analyzer's kit.

Thermal printer connection option – to be chosen by the client.

The milk analyzer has 3 calibration channels, which can be set for 3 types of milk, chosen by the client.

One calibration is included in the analyzer's price – the other/s should be paid.

Cleaning – automatic, with the **MilkoScope – Flush** cleaning concentrate.

Written and sound message for necessity of cleaning.

Suitable for all kinds of animal milk



Scope Electric Bulgaria
Razgrad

Phone/Fax: +359-84-660-654

Mobile: +359-888-89-10-89

E-mail: office@scope-electric.com

www.scope-electric.com

Scope Electric Germany
Regensburg

Phone: +49-941-690-7945

Mobile: +49-1797532699

E-mail: office.de@scope-electric.com

www.scope-electric.com



Scope Electric

Manufacturer of professional milk analyzers

MilkoScope

Julie C3 Automatic

Measuring Ranges:		Technical Overview:	
Fat:	0% - 15%	Sample Volume:	5-12 ml.
Protein:	0% - 8%	Sample Speed:	65 sec.
Lactose:	0% - 7%	Sample Temperature:	5 - 30°C
Solids-Non-Fat (SNF):	0% - 15%	Humidity:	0 - 80% RH
Solids:	0% - 5%	Database Processing:	RS 232 port
Density:	1000-1200 kg/m3	Power Supply:	110/230VAC $\pm$ 10%,50/60Hz
Added Water:	0% - 100%	Power Consumption:	30W
Freezing Point:	-0,400 to -0,700°C	Dimension:	H=280 mm,W=170 mm,D=245 mm
Sample Temperature:	5 - 30°C	Weight:	3,1 kg

Accuracy:			Kit:	
Fat:	$\pm 0.04\%$	against Rose Gottlieb	Milk Analyzer	
Protein:	$\pm 0.03\%$	against Kjeldahl	Power Supply Cable 220V	
Lactose:	$\pm 0.03\%$	against Polarimetry	Serial Cable to PC	1 piece
Solids-Non-Fat (SNF):	$\pm 0.05\%$	against Drying Oven	Sample Cup	1 piece
Solids:	$\pm 0.03\%$	against Gravimetric	Waste Container	1 piece
Density:	$\pm 0,3$ kg/m3	against pycnometer	Spare Hose	1 piece
Added Water:	$\pm 2\%$	against cryoscope	Cleaning Concentrate	1 piece
Freezing Point:	0,005 °C		User's Manual	1 piece
			CD with software	

Deviation at numerous analyses of one and the same sample: $\pm 0.02\%$ for all parameters.

Scope Electric Bulgaria
Razgrad
Phone/Fax: +359-84-660-654
Mobile: +359-888-89-10-89
E-mail: office@scope-electric.com
www.scope-electric.com

Scope Electric Germany
Regensburg
Phone: +49-941-690-7945
Mobile: +49-1797532699
E-mail: office.de@scope-electric.com
www.scope-electric.com